

Calibration Certificate

Accredited Calibration



Cal Cert ID: 161006

Page 1 of 2

Cust. Name: WASHINGTON STATE PATROL
Address: 222 TUMWATER BLVD.
 BUILDING 16
City: TUMWATER
State/Zip: WA 98504 US

Instr SN: 410004SY/651314	Out1 Min Range: 8.0000 psi	Cal Date: 01/22/2019
Instr Descr: CPG 2400	Out1 Max Range: 17.0000 psi	Cal Time: 8:51 am
Sensor SN: 410004SY/651314	Limit of Error: 0.000 %F.S.+ 0.030 % R	Cal Tech: SALLYG
Procedure ID: WI02063	Pressure Type: Absolute	Order Nbr: 26784

Test Points Values and Readings

Test Point #	Reference Value PSI	Measured Values psi	Error % R	Uncertainty psi
1	8.00406	8.0041	0.000	.00018
2	9.80394	9.8040	0.001	.00022
3	11.60379	11.6040	0.002	.00026
4	13.40370	13.4038	0.001	.00029
5	15.20358	15.2036	0.000	.00033
6	17.00351	17.0034	-0.001	.00037
7	15.20355	15.2036	0.000	.00033
8	13.40370	13.4038	0.001	.00029
9	11.60381	11.6039	0.001	.00026
10	9.80392	9.8040	0.001	.00022
11	8.00406	8.0041	0.001	.00018

Measurement Standards Used in Calibration

Std Type	Serial Number	Description	Assoc. Test Reports	Recal Date
CPD 8000	10017/10016/9994	DH CPD 8000	CAL ID: 159795	12/11/2021

The combined and expanded uncertainty is reported at K=2 providing an approximate 95% confidence level.

Remarks and Notes

As Found Calibration.
 As Left Calibration.
 Prior to "AS FOUND" calibration, zero was adjusted +0.0026 PSI
 Recalibration Date: 01/22/2020
 An asterisk (*) denotes a measured condition where the error exceeds the maximum permissible error.
 Dash numbers preceded by the sensor SN reference a primary (-1) or turndown (-2) range as applicable.
 The standards and calibration program at Mensor complies with the requirements of ISO/IEC 17025:2005, ANSI/NCSL Z540-1 and Mensor Management Manual MM 3.0, Revision A, dated June 6, 2018.
 Standards used in this calibration are traceable to SI units of measurements through N.I.S.T. or recognized national metrology institutes. All calibrations are performed in ambient conditions of 23 +/- 4 deg C and 20-80% relative humidity.
 This calibration certificate shall not be reproduced, except in full, without the written approval of Mensor.



Cal Cert ID: 161006

Page 2 of 2

I, Sally Guerrero, certify the accuracy of this Calibration Certificate:

Calibration Lab Technician

Mensor Calibration Certificates and Stickers

Mensor calibration certificates and stickers do not contain calibration intervals unless specified by the customer, as Mensor will not dictate to our customers when they should re-calibrate their instruments.

The Calibration Sticker that is with this product is for your records. Due to the various location requirements of our customers, we do not affix the Calibration Sticker to the product.

For Mensor products, refer to the Calibration Section of your operation manual for the recommended calibration interval.

Stramer
1-30-2019