GUTH LABORATORIES, INC.

REPORT OF CALIBRATION
GUTH-EUTECHNICS DIGITAL THERMOMETER
29.5° C – 38.5° C

GUTH P/N GU-4300-DT

S/N 091796

RECALIBRATION DATE 3-30-01

The Digital Thermometer calibration was verified at 30.00° C, 34.00° C, and 37.00° C by comparison to temperature standards traceable to the National Institute of Standards and Technology.

BATH TEMPERATURE (°C)	MASTER DIGITAL THERMOMETER S/N 074857	DIGITAL THERMOMETER READING	TEST DATE
30.00	30.00	30.00	
34.00	34.00	34.00	
37.00	37.00	37.00	

Tested by D.A.N.

Date 3-3-00

Approved_____

Date_____