

# WASHINGTON STATE PATROL

## Report of Calibration

Guth-Eutechnics Digital Thermometer

Guth Part Number: GU-4300-DT

Serial Number: 091796

The Digital Thermometer calibration was verified at 30.00°C, 34.00°C, and 37.00°C by comparison to temperature standards traceable to the National Institute of Standards & Technology.

Simulator Temperature Degrees C	Master Digital Thermometer S/N 082708	Digital Thermometer Reading	Test Date
30.00	30.00°C	30.00°C	09/11/00
34.00	34.00°C	34.00°C	09/11/00
37.00	37.00°C	37.00°C	09/11/00

Tested by: TRAB

Date: 9-11-00

Bostec Inc.