



Certificate of Calibration

Everett Service Center

Certificate Number: 649092-69352611:1348673545

Result Summary:	PASS	Date of Calibration:	26 September 2012
Data Type:	FOUND-LEFT	Recommended Due Date:	26 September 2013
Manufacturer:	FLUKE	Date of Certificate:	28 September 2012
Model:	73 III	Received Date:	25 September 2012
Serial Number:	69352611	Temperature:	23.8 °C
Description:	MULTIMETER	Relative Humidity:	38 %RH
Procedure Name:	FLUKE 73-3: (1 YEAR) CAL VER/ALT 5520A	Procedure Revision:	1.0

Customer Name:	STATE OF WASHINGTON
City, State:	SEATTLE, WA
Customer Item ID:	DESTITO# 435
PO Number:	IDS 9/24/12
RMA Number:	30125310

This calibration is traceable to the International System of Units (SI), through National Metrology Institutes, radiometric techniques, or natural physical constants. This certificate applies only to the item identified and shall not be reproduced other than in full, without the specific written approval by Fluke Corporation. Calibration certificates without signature are not valid. The calibration has been completed in accordance with Fluke Electronics Corporation Quality System Document 111.0 Rev 114 08/09 and Fluke Customer Support Services QAM 400 Rev. 002 03/22/2012.

The Data Type found in this certificate must be interpreted as:

- As - Found Calibration data collected before the unit is adjusted and / or repaired.
- As - Left Calibration data collected after the unit has been adjusted and / or repaired.
- Found-Left Calibration data collected without any adjustment and / or repair performed.

Corey Dinardi
Metrology Technician



Certificate Number:
649092-69352611:1348673545

Calibration Date:
26-Sep-12

Standards Used

Asset #	Instrument Model	Cal Date	Cal Due
15267	FLUKE 5520A/1GHZ 120 CALIBRATOR	17 February 2012	17 February 2013

End of Report