



Certificate of Calibration Everett Service Center

ISO 9001:2008 (10100/2)

Certificate Number: 649092-69352611:1377042573

Result Summary: PASS	Date of Calibration: 20 August 2013
Data Type: AS-LEFT	Recommended Due Date: 20 August 2014
Manufacturer: FLUKE	Date of Certificate: 21 August 2013
Model: 73 III	Received Date: 16 August 2013
Serial Number: 69352611	Temperature: 22.9 °C
Description: MULTIMETER	Relative Humidity: 42 %RH
Procedure Name: FLUKE 73-3: (1 YEAR) CAL VER/ALT 5520A	Procedure Revision: 1.0

Customer Name: STATE OF WASHINGTON
City, State: SEATTLE, WA
Customer Item ID: 69352611
PO Number: WSPCCS
RMA Number: 30349853

This calibration is traceable to the International System of Units (SI), through National Metrology Institutes, radiometric techniques, or natural physical constants. This certificate applies only to the item identified and shall not be reproduced other than in full, without the specific written approval by Fluke Corporation. Calibration certificates without signature are not valid. The calibration has been completed in accordance with Fluke Electronics Corporation Quality System Document 111.0 Rev 116 08/12 and Fluke Customer Support Services QAM 400 Rev. 002 03/22/2012.

The Data Type found in this certificate must be interpreted as:

- As - Found Calibration data collected before the unit is adjusted and / or repaired.
- As - Left Calibration data collected after the unit has been adjusted and / or repaired.
- Found-Left Calibration data collected without any adjustment and / or repair performed.

Corey Dinardi
Metrology Technician

FLUKE ®

Certificate Number:
649092-69352611:1377042573

Calibration Date:
20-Aug-13

Standards Used

Asset #	Instrument Model	Cal Date	Cal Due
14827	FLUKE 5520A CALIBRATOR	11 April 2013	11 October 2013

End of Report