



## Certificate of Calibration

### Everett Service Center

**Certificate Number:** 649084-69920523:1359110064

<b>Result Summary:</b> PASS	<b>Date of Calibration:</b> 25 January 2013
<b>Data Type:</b> FOUND-LEFT	<b>Recommended Due Date:</b> 25 January 2014
<b>Manufacturer:</b> FLUKE	<b>Date of Certificate:</b> 25 January 2013
<b>Model:</b> 70 III	<b>Received Date:</b> 25 January 2013
<b>Serial Number:</b> 69920523	<b>Temperature:</b> 23.7 °C
<b>Description:</b> MULTIMETER	<b>Relative Humidity:</b> 33 %RH
<b>Procedure Name:</b> Fluke 70-3: (1 year) CAL VER/Alt 5520A	<b>Procedure Revision:</b> 1.0

**Customer Name:** STATE OF WASHINGTON  
**City, State:** SEATTLE, WA  
**Customer Item ID:** 69920523  
**PO Number:** CCS 23-JAN-2013  
**RMA Number:** 30206614

This calibration is traceable to the International System of Units (SI), through National Metrology Institutes, radiometric techniques, or natural physical constants. This certificate applies only to the item identified and shall not be reproduced other than in full, without the specific written approval by Fluke Corporation. Calibration certificates without signature are not valid. The calibration has been completed in accordance with Fluke Electronics Corporation Quality System Document 111.0 Rev 116 08/12 and Fluke Customer Support Services QAM 400 Rev. 002 03/22/2012.

The Data Type found in this certificate must be interpreted as:

- As - Found Calibration data collected before the unit is adjusted and / or repaired.
- As - Left Calibration data collected after the unit has been adjusted and / or repaired.
- Found-Left Calibration data collected without any adjustment and / or repair performed.

Rocio Carrillo  
Metrology Technician

**FLUKE**®

Certificate Number:  
649084-69920523:1359110064

Calibration Date:  
25-Jan-13

**Standards Used**

Asset #	Instrument Model	Cal Date	Cal Due
13607	FLUKE 5520A CALIBRATOR	21 January 2013	21 January 2014

End of Report