

Certificate of Calibration Everett Service Center



ISO 9001:2008 (10100/2)

Certificate Number: 649084-70500451:1352743133

Result Summary:	PASS	Date of Calibration:	12 November 2012
Data Type:	FOUND-LEFT	Recommended Due Date:	12 November 2013
Manufacturer:	FLUKE	Date of Certificate:	12 November 2012
Model:	70 III	Received Date:	12 November 2012
Serial Number:	70500451	Temperature:	22.8 °C
Description:	MULTIMETER	Relative Humidity:	36 %RH
Procedure Name:	FLUKE 70-3: (1 YEAR) CAL VER/ALT 5520A	Procedure Revision:	1.0

Customer Name: STATE OF WASHINGTON
City, State: SEATTLE, WA
Customer Item ID: 70500451
PO Number: CCS WASHINGTON STATE PATROL
RMA Number: 30158779

This calibration is traceable to the International System of Units (SI), through National Metrology Institutes, radiometric techniques, or natural physical constants. This certificate applies only to the item identified and shall not be reproduced other than in full, without the specific written approval by Fluke Corporation. Calibration certificates without signature are not valid. The calibration has been completed in accordance with Fluke Electronics Corporation Quality System Document 111.0 Rev 116 08/12 and Fluke Customer Support Services QAM 400 Rev. 002 03/22/2012.

The Data Type found in this certificate must be interpreted as:

- As - Found Calibration data collected before the unit is adjusted and / or repaired.
- As - Left Calibration data collected after the unit has been adjusted and / or repaired.
- Found-Left Calibration data collected without any adjustment and / or repair performed.

A handwritten signature in black ink, appearing to read 'Corey Dinardi'.

Corey Dinardi
Metrology Technician

FLUKE®

Certificate Number:
649084-70500451:1352743133

Calibration Date:
12-Nov-12

Standards Used

Asset #	Instrument Model	Cal Date	Cal Due
11733	FLUKE 5520A/3 120 CALIBRATOR	09 April 2012	09 January 2013

End of Report