



## Certificate of Calibration

### Everett Service Center

ISO 9001:2008 (10100/2)

Certificate Number: 649084-70530846:1348685502

Result Summary:	PASS	Date of Calibration:	26 September 2012
Data Type:	FOUND-LEFT	Recommended Due Date:	26 September 2013
Manufacturer:	FLUKE	Date of Certificate:	26 September 2012
Model:	70 III	Received Date:	26 September 2012
Serial Number:	70530846	Temperature:	24.0 °C
Description:	MULTIMETER	Relative Humidity:	42 %RH
Procedure Name:	FLUKE 70-3: (1 YEAR) CAL VER/5500	Procedure Revision:	2.0

Customer Name: STATE OF WASHINGTON  
City, State: SEATTLE, WA  
Customer Item ID: 70530846  
PO Number: IDS CCS 20-SEP-2012  
RMA Number: 30123182

This calibration is traceable to the International System of Units (SI), through National Metrology Institutes, radiometric techniques, or natural physical constants. This certificate applies only to the item identified and shall not be reproduced other than in full, without the specific written approval by Fluke Corporation. Calibration certificates without signature are not valid. The calibration has been completed in accordance with Fluke Electronics Corporation Quality System Document 111.0 Rev 114 08/09 and Fluke Customer Support Services QAM 400 Rev. 002 03/22/2012.

The Data Type found in this certificate must be interpreted as:

- As - Found Calibration data collected before the unit is adjusted and / or repaired.
- As - Left Calibration data collected after the unit has been adjusted and / or repaired.
- Found-Left Calibration data collected without any adjustment and / or repair performed.

Long Le  
Metrology Technician

**FLUKE**®

Certificate Number:  
649084-70530846:1348685502

Calibration Date:  
26-Sep-12

**Standards Used**

Asset #	Instrument Model	Cal Date	Cal Due
10113	FLUKE 5500A CALIBRATOR	30 January 2012	30 January 2013

End of Report