



ISO 9001:2008 (10100/2)

Certificate of Calibration Everett Service Center

Certificate Number: 649084-70530846:1380724374

Result Summary: PASS	Date of Calibration: 02 October 2013
Data Type: FOUND-LEFT	Recommended Due Date: 02 October 2014
Manufacturer: FLUKE	Date of Certificate: 02 October 2013
Model: 70 III	Received Date: 24 September 2013
Serial Number: 70530846	Temperature: 23.0 °C
Description: MULTIMETER	Relative Humidity: 40 %RH
Procedure Name: Fluke 70-3: (1 year) CAL VER/Ait 5520A	Procedure Revision: 1.0

Customer Name: STATE OF WASHINGTON
City, State: SEATTLE, WA
Customer Item ID:
PO Number: WSP-CCS
RMA Number: 30376264

This calibration is traceable to the International System of Units (SI), through National Metrology Institutes, radiometric techniques, or natural physical constants. This certificate applies only to the item identified and shall not be reproduced other than in full, without the specific written approval by Fluke Corporation. Calibration certificates without signature are not valid. The calibration has been completed in accordance with Fluke Electronics Corporation Quality System Document 111.0 Rev 116 08/12 and Fluke Customer Support Services QAM 400 Rev. 002 03/22/2012.

The Data Type found in this certificate must be interpreted as:

- As - Found Calibration data collected before the unit is adjusted and / or repaired.
- As - Left Calibration data collected after the unit has been adjusted and / or repaired.
- Found-Left Calibration data collected without any adjustment and / or repair performed.


Rocio Carrillo
Metrology Technician

FLUKE ®

Certificate Number:
649084-70530846:1380724374

Calibration Date:
2-Oct-13

Standards Used

Asset #	Instrument Model	Cal Date	Cal Due
13682	FLUKE 5520A CALIBRATOR	16 August 2013	16 November 2013

End of Report