



## Certificate of Calibration

Everett Service Center

Certificate Number: 649084-70530850:1352207939

<b>Result Summary:</b> PASS	<b>Date of Calibration:</b> 06 November 2012
<b>Data Type:</b> FOUND-LEFT	<b>Recommended Due Date:</b> 06 November 2013
<b>Manufacturer:</b> FLUKE	<b>Date of Certificate:</b> 06 November 2012
<b>Model:</b> 70 III	<b>Received Date:</b> 05 November 2012
<b>Serial Number:</b> 70530850	<b>Temperature:</b> 24.2 °C
<b>Description:</b> MULTIMETER	<b>Relative Humidity:</b> 38 %RH
<b>Procedure Name:</b> FLUKE 70-3: (1 YEAR) CAL VER/ALT 5520A	<b>Procedure Revision:</b> 1.0

**Customer Name:** STATE OF WASHINGTON  
**City, State:** SEATTLE, WA  
**Customer Item ID:** 70530850  
**PO Number:** CCS  
**RMA Number:** 30154152

This calibration is traceable to the International System of Units (SI), through National Metrology Institutes, radiometric techniques, or natural physical constants. This certificate applies only to the item identified and shall not be reproduced other than in full, without the specific written approval by Fluke Corporation. Calibration certificates without signature are not valid. The calibration has been completed in accordance with Fluke Electronics Corporation Quality System Document 111.0 Rev 116 08/12 and Fluke Customer Support Services QAM 400 Rev. 002 03/22/2012.

The Data Type found in this certificate must be interpreted as:

- As - Found Calibration data collected before the unit is adjusted and / or repaired.
- As - Left Calibration data collected after the unit has been adjusted and / or repaired.
- Found-Left Calibration data collected without any adjustment and / or repair performed.

  
Rocio Carrillo  
Metrology Technician

**FLUKE** ®

Certificate Number:  
649084-70530850:1352207939

Calibration Date:  
6-Nov-12

**Standards Used**

Asset #	Instrument Model	Cal Date	Cal Due
J1670	FLUKE 5520A CALIBRATOR	09 July 2012	09 July 2013

End of Report