



## Certificate of Calibration Everett Service Center

ISO 9001:2008 (10100/2)

Certificate Number: 649084-708600863:1341252198

Result Summary:	PASS	Date of Calibration:	02 July 2012
Data Type:	FOUND-LEFT	Recommended Due Date:	02 July 2013
Manufacturer:	FLUKE	Date of Certificate:	02 July 2012
Model:	70 III	Received Date:	29 June 2012
Serial Number:	708600863	Temperature:	24.0 °C
Description:	MULTIMETER	Relative Humidity:	39 %RH
Procedure Name:	FLUKE 70-3: (1 YEAR) CAL VER/ALT 5520A	Procedure Revision:	1.0

Customer Name: STATE OF WASHINGTON  
City, State: SEATTLE, WA  
Customer Item ID: 708600863  
PO Number: IDS  
RMA Number: 30063295

This calibration is traceable to the International System of Units (SI), through National Metrology Institutes, radiometric techniques, or natural physical constants. This certificate applies only to the item identified and shall not be reproduced other than in full, without the specific written approval by Fluke Corporation. Calibration certificates without signature are not valid. The calibration has been completed in accordance with Fluke Electronics Corporation Quality System Document 111.0 Rev 114 08/09 and Fluke Customer Support Services QAM 400 Rev. 002 03/22/2012.

The Data Type found in this certificate must be interpreted as:

- As - Found Calibration data collected before the unit is adjusted and / or repaired.
- As - Left Calibration data collected after the unit has been adjusted and / or repaired.
- Found-Left Calibration data collected without any adjustment and / or repair performed.

A handwritten signature in black ink, appearing to read 'Corey Dinardi'.

Corey Dinardi  
Metrology Technician

**Standards Used**

<b>Asset #</b>	<b>Instrument Model</b>	<b>Cal Date</b>	<b>Cal Due</b>
15261	FLUKE 5520A/1GHZ CALIBRATOR	16 April 2012	16 October 2012

End of Report