



Certificate of Calibration Everett Service Center

ISO 9001:2008 (10100/2)

Certificate Number: 649084-708601073:1376042758

Result Summary: PASS	Date of Calibration: 09 August 2013
Data Type: FOUND-LEFT	Recommended Due Date: 09 August 2014
Manufacturer: FLUKE	Date of Certificate: 09 August 2013
Model: 70 III	Received Date: 02 August 2013
Serial Number: 708601073	Temperature: 22.9 °C
Description: MULTIMETER	Relative Humidity: 46 %RH
Procedure Name: FLUKE 70-3: (1 YEAR) CAL VER/ALT 5520A	Procedure Revision: 1.0

Customer Name: STATE OF WASHINGTON
City, State: SEATTLE, WA
Customer Item ID: MADSEN
PO Number: CCS IMPAIRED DRIVING 5 WASHINGTON STATE PATROL
RMA Number: 30340009

This calibration is traceable to the International System of Units (SI), through National Metrology Institutes, radiometric techniques, or natural physical constants. This certificate applies only to the item identified and shall not be reproduced other than in full, without the specific written approval by Fluke Corporation. Calibration certificates without signature are not valid. The calibration has been completed in accordance with Fluke Electronics Corporation Quality System Document 111.0 Rev 116 08/12 and Fluke Customer Support Services QAM 400 Rev. 002 03/22/2012.

The Data Type found in this certificate must be interpreted as:

- As - Found Calibration data collected before the unit is adjusted and / or repaired.
- As - Left Calibration data collected after the unit has been adjusted and / or repaired.
- Found-Left Calibration data collected without any adjustment and / or repair performed.

William Nelson
Metrology Technician

FLUKE ®

Certificate Number:
649084-708601073:1376042758

Calibration Date:
9-Aug-13

Standards Used

Asset #	Instrument Model	Cal Date	Cal Due
11733	FLUKE 5520A/3 120 CALIBRATOR	01 August 2013	01 February 2014

End of Report