

# Z540.1

**FLUKE**

## Certificate of Calibration

### Beaverton Service Center

**Certificate Number:** 88447

**Data Type:** Found-Left  
**Result Summary:** In Tolerance  
**Manufacturer:** Fluke  
**Model:** 70 III  
**Serial Number:** 708601073  
**Description:** Multimeter

**Calibration Date:** 28-Nov-2014  
**Calibration Due:** 28-Nov-2015  
**Certificate Date:** 28-Nov-2014  
**Temperature:** 22.8 °C  
**Humidity:** 46.9 %

**Procedure:** Fluke 70-3: (1 year) CAL VER /5520A**Revision:** 1.2**Customer:** STATE OF WASHINGTON**City:** SEATTLE**Country:** US**State:** WA**Purchase Order:** CCS WSP**RMA:** 30639910

This calibration is traceable to the International System of Units (SI), through National Metrology Institutes (NIST, PTB, NRC, NPL, etc.), radiometric techniques, or natural physical constants. This certificate applies only to the item identified and shall not be reproduced other than in full, without the specific written approval by Fluke Corporation. Calibration certificates without signature are not valid. The calibration has been completed in accordance with Fluke Electronics Corporation Quality System Document 111.0, Fluke 17025 Quality Manual QSD 111.41 and/or Fluke Customer Support Services QAM 400.

The Data Type found in this certificate must be interpreted as:

- As - Found Calibration data collected before the unit is adjusted and / or repaired.
- As - Left Calibration data collected after the unit has been adjusted and / or repaired.
- Found-Left Calibration data collected without any adjustment and / or repair performed.

This calibration conforms to the requirements of ANSI/NCCL Z540-1-1994 (R2002).


In the attached measurement results, deviation may be expressed with units, Measured Value (MV) - Nominal Value (NV) or as a proportion of the nominal value  $((MV-NV)/NV)$ , expressed without units with a scalar multiplier such as % (0.01), or as a ratio of the units (mA/A,  $\mu V/V$ , etc.) Descriptions such as  $\mu A/A$ ,  $\mu V/V$ , and others, where used to annotate results or column headings are the preferred replacements for what was historically labeled as "ppm" or parts-per-million and described the results in that column, unless otherwise noted by units symbols.


Where applicable, the expanded uncertainty of measurement at the time of test is given in the following pages. They are calculated in accordance with the method described in the ISO Guide to the Expression of Uncertainty in Measurement (GUM). The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k, such that the confidence level approximates 95%.

Where applicable, the Test Uncertainty Ratio (TUR) is provided in the following pages. Unless otherwise stated, the TUR for a given measurement result is 4:1 or greater.

Results are reviewed to establish where any measurement results exceeded the manufacturer's specifications.

Measurement results greater than limits of error are indicated by "I".



  
MICHAEL GRIFFITHS  
Issued By

**Certificate Number:** 88447**Date of Calibration:** 28-Nov-2014**Standards Used**

<b>Asset</b>	<b>Description</b>	<b>Cal-Date</b>	<b>Cal-Due</b>
16533	Fluke 5522A Calibrator	12-Jul-2014	12-Jul-2015

Certificate Number: 88447

Date of Calibration: 28-Nov-2014

**Calibration Data**

Parameter	Nominal Value	Measurement Result	Limits of Error		Test Uncertainty Ratio (TUR)
			Lower Limit	Upper Limit	
<b>DISPLAY TEST</b>					
Display Segments Illuminated Correctly:		Pass			
<b>RESISTANCE TEST</b>					
<b>320Ohms Range</b>					
0.0 Ohm	0.00	0.0	0.0	0.2	
100.0 Ohm	100.00	99.9	99.3	100.7	
<b>3200Ohms Range</b>					
1000 Ohm	1000.0	999	994	1006	
<b>32kOhms Range</b>					
10.00 kOhm	10.000	10.00	9.94	10.06	
<b>320kOhms Range</b>					
100.0 kOhm	100.00	100.0	99.4	100.6	
<b>3.2MOhms Range</b>					
1.000 MOhm	1.0000	1.000	0.994	1.006	
<b>32MOhms Range</b>					
10.00 MOhm	10.000	10.00	9.79	10.21	
<b>DIODE TEST</b>					
Beeper Audible		Pass			
Beeper OFF		Pass			
<b>DC VOLTAGE TEST</b>					
<b>320mV Range</b>					
300.0 mV	300.00	299.8	299.0	301.0	
<b>DC VOLTAGE TEST</b>					
<b>3.2V Range</b>					
2.700 V	2.7000	2.700	2.691	2.709	
<b>32V Range</b>					
27.00 V	27.000	27.00	26.91	27.09	
<b>320V Range</b>					
270.0 V	270.00	270.0	269.1	270.9	
<b>600V Range</b>					
600 V	600.0	601	597	603	
<b>AC VOLTAGE TEST</b>					
<b>3.2V Range</b>					
2.700 V @ 100 Hz	2.7000	2.700	2.644	2.756	
2.700 V @ 500 Hz	2.7000	2.700	2.644	2.756	

Certificate Number: 88447

Date of Calibration: 28-Nov-2014

## Calibration Data

Parameter	Nominal Value	Measurement Result	Limits of Error		Test Uncertainty Ratio (TUR)
			Lower Limit	Upper Limit	
<b>600V Range</b>					
600 V @ 100 Hz	600.0	601	586	614	
600 V @ 1 kHz	600.0	600	586	614	