



Calibration Certificate

Description:	MULTIMETER	Certificate Number:	649084-73700066:1217837775
Manufacturer:	FLUKE	Date of Calibration:	04 August 2008
Model:	70 III	Date of Certificate:	04 August 2008
Serial Number:	73700066	Date Due:	04 August 2009
Customer Name:	STATE OF WASHINGTON	Procedure Name:	MFG MANUAL
City, State:	SEATTLE, WA	Procedure Revision:	7/98
Customer Item ID:	73700066	Data Type:	FOUND-LEFT
PO Number:	SATO CCS	Temperature:	23 ± 3.0 °Celsius
RMA Number:	4021486	Relative Humidity:	25% ≤ RH ≤ 60%
		Test Result:	PASS

The Fluke Corporation, NQA ISO 9001:2000 ISO Certification No. 10100/2, certifies that the instrument identified above was calibrated in accordance with applicable Fluke calibration procedures. Its calibration processes are ISO-9001 controlled and are designed to certify that the instrument was within its published specifications at the time of calibration.

The measurement standards and instruments used during the calibration of this instrument are traceable to the United States National Institute of Standards and Technology (NIST), other reputable National Institutes, natural physical constants, consensus standards, or by ratio type measurements.

This certificate applies to only the item identified and shall not be reproduced other than in full, without the specific written approval by Fluke Corporation. The user is obliged to have the in object recalibrated at appropriate intervals. Calibration Certificates without signature are not valid.

The Data type that could be found in this certificate is interpreted as follows:

- As Found — The unit needed adjustment and/or repair.
- As Left — The unit was adjusted and/or repaired.
- As Found/ As Left — The unit was calibrated without any adjustment and/or repair performed.

Comments:

Long Le
Metrology Technician

Standards Used

Asset #	Instrument Model	Cal Date	Cal Due
10127	FLUKE 5520A CALIBRATOR	28 November 2007	28 August 2008

End of Report