

1420 75th St. SW
Everett, Washington 98203
USA

NQA ISO 9001:2000 (10100/2)

Calibration Certificate

Description:	MULTIMETER	Certificate Number:	649084-73700066:1273223263
Manufacturer:	FLUKE	Date of Calibration:	07 May 2010
Model:	70 III	Date of Certificate:	07 May 2010
Serial Number:	73700066	Recommended Due Date:	07 May 2011
Customer Name:		Procedure Name:	
STATE OF WASHINGTON		MFG MANUAL	
City, State:	SEATTLE, WA	Procedure Revision:	7/98
Customer Item ID:	73700066	Data Type:	FOUND-LEFT
PO Number:	SATO CCS 05/O06/O10	Temperature:	23.00 °Celsius
RMA Number:	4464335	Relative Humidity:	40 %
Result Summary:	PASS		
Received Date:			

The Data type that could be found in this certificate must be interpreted as:

- As-Found - Calibration data collected before the unit is adjusted and/or repaired.
- As-Left - Calibration data collected after the unit is adjusted and/or repaired.
- Found-Left - Calibration data collected without any adjustment and/or repair performed.

This certificate applies only to the item identified and shall not be reproduced other than in full, without the specific written approval by Fluke Corporation. The user is obliged to have the object recalibrated at appropriate intervals.

Comments:Long Le
Metrology Technician

Traceability Information

For each parameter listed below the calibration was conducted using an unbroken chain of standards to:

DC VoltageThe Voltage Reference standard group, traceable to the Fluke Primary Standards Laboratory, which is traceable to the U.S. representation of the volt, through the internationally accepted value of the Josephson constant $K_j=483597.9$ GHz/V and a 10 Volt Josephson Array Voltage Standard.**Frequency and Period**

The GPS-Rubidium Disciplined oscillator frequency standard, traceable to the United States Naval Observatory (USNO), which is traceable to the National Institute of Standards and Technology.

AC Voltage, Resistance, DC Current, AC Current, Capacitance, Inductance, Phase

The Fluke Primary Standards Laboratory, which is traceable to the National Institute of Standards and Technology.

AC Voltage Flatness

The Fluke Primary Standards Laboratory, or Agilent Technologies Standards Laboratory which are traceable to the National Institute of Standards and Technology.

Humidity

The Vaisala Measurement Standards Laboratory Primary Salt calibration bath, with traceability based on the physical phenomena in which the equilibrium relative humidity values associated with certain saturated salt solutions are known.

Rise Time

The Tektronix GmbH Calibration Laboratory which is traceable to the Physikalisch-Technische Bundesanstalt.

Radiation Temperature

The National Institute of Standards and Technology, the Physikalisch-Technische Bundesanstalt, or Hart Scientific.

Contact Temperature

The Fluke Primary Standards Laboratory, Hart Scientific, which are traceable to the National Institute of Standards and Technology.

Gas Flow

The DHI Calibration Laboratory, which is traceable to the National Institute of Standards and Technology.

Pressure

The DHI Calibration Laboratory, which is traceable to the Laboratoire National D'Essais, Physikalisch-Technische Bundesanstalt and National Institute of Standards and Technology, or traceable to the Mensor or Ashcroft Calibration Laboratories, which are traceable to the National Institute of Standards and Technology.

Standards Used

Asset #	Instrument Model	Cal Date	Cal Due
10127	FLUKE 5520A CALIBRATOR	29 May 2009	29 May 2010

End of Report