

Certificate of Calibration Everett Service Center



ISO 9001:2008 (10100/2)

Certificate Number: 649084-73700066:1340966812

Result Summary: PASS	Date of Calibration: 29 June 2012
Data Type: FOUND-LEFT	Recommended Due Date: 29 June 2013
Manufacturer: FLUKE	Date of Certificate: 29 June 2012
Model: 70 III	Received Date: 28 June 2012
Serial Number: 73700066	Temperature: 23.9 °C
Description: MULTIMETER	Relative Humidity: 45 %RH
Procedure Name: FLUKE 70-3: (1 YEAR) CAL VER/ALT 5520A	Procedure Revision: 1.0

Customer Name: STATE OF WASHINGTON
City, State: SEATTLE, WA
Customer Item ID: 73700066
PO Number: CCS IMPAIRED DRIVING
RMA Number: 30062875

This calibration is traceable to the International System of Units (SI), through National Metrology Institutes, radiometric techniques, or natural physical constants. This certificate applies only to the item identified and shall not be reproduced other than in full, without the specific written approval by Fluke Corporation. Calibration certificates without signature are not valid. The calibration has been completed in accordance with Fluke Electronics Corporation Quality System Document 111.0 Rev 114 08/09 and Fluke Customer Support Services QAM 400 Rev. 002 03/22/2012.

The Data Type found in this certificate must be interpreted as:

- As - Found Calibration data collected before the unit is adjusted and / or repaired.
- As - Left Calibration data collected after the unit has been adjusted and / or repaired.
- Found-Left Calibration data collected without any adjustment and / or repair performed.

Charles Oestreich
Metrology Technician

FLUKE ®

Certificate Number:
649084-73700066:1340966812

Calibration Date:
29-Jun-12

Standards Used

Asset #	Instrument Model	Cal Date	Cal Due
15267	FLUKE 5520A/1GHZ 120 CALIBRATOR	17 February 2012	17 February 2013

End of Report