



Certificate of Calibration

Everett Service Center

ISO 9001:2008 (10100/2)

Certificate Number: 649084-75230619:1368046480

Result Summary: PASS	Date of Calibration: 08 May 2013
Data Type: FOUND-LEFT	Recommended Due Date: 08 May 2014
Manufacturer: FLUKE	Date of Certificate: 08 May 2013
Model: 70 III	Received Date: 02 May 2013
Serial Number: 75230619	Temperature: 22.4 °C
Description: MULTIMETER	Relative Humidity: 38 %RH
Procedure Name: FLUKE 70-3: (1 YEAR) CAL VER/ALT 5520A	Procedure Revision: 1.0

Customer Name: STATE OF WASHINGTON
City, State: SEATTLE, WA
Customer Item ID: 75230619
PO Number: CCS 01-MAY-2013
RMA Number: 30277384

This calibration is traceable to the International System of Units (SI), through National Metrology Institutes, radiometric techniques, or natural physical constants. This certificate applies only to the item identified and shall not be reproduced other than in full, without the specific written approval by Fluke Corporation. Calibration certificates without signature are not valid. The calibration has been completed in accordance with Fluke Electronics Corporation Quality System Document 111.0 Rev 116 08/12 and Fluke Customer Support Services QAM 400 Rev. 002 03/22/2012.

The Data Type found in this certificate must be interpreted as:

- As - Found Calibration data collected before the unit is adjusted and / or repaired.
- As - Left Calibration data collected after the unit has been adjusted and / or repaired.
- Found-Left Calibration data collected without any adjustment and / or repair performed.

A handwritten signature in black ink, appearing to read 'Corey Dinardi'.

Corey Dinardi
Metrology Technician

FLUKE®Certificate Number:
649084-75230619:1368046480Calibration Date:
8-May-13**Standards Used**

Asset #	Instrument Model	Cal Date	Cal Due
15267	FLUKE 5520A/1GHZ 120 CALIBRATOR	27 February 2013	27 February 2014

End of Report