


**INTEROFFICE COMMUNICATION**  
**WASHINGTON STATE PATROL**



**TO:** Sergeant Rod Gullberg, Breath Test Program  
**FROM:** Dr. Barry K. Logan, Forensic Laboratory Services Bureau  
**SUBJECT:** CDM Approval  
**DATE:** December 20, 2004

I have reviewed the QAP testing data from 11/30/2004 performed on a series of five CDM DataMaster instruments (serial numbers 140002, 140021, 140010, 140013 and 140011). All instruments were tested with software version 76043-004 (04/28/04) installed.

The instruments performed exceedingly well, meeting the standards of the QAP procedure in all respects.

Through this notice I am approving the Software version 76043-004 (04/28/04) for use in field instruments for evidential breath testing.

I also approve the deployment of DataMaster CDM instruments to the field for evidential breath testing. Each instrument deployed to the field must successfully pass the QAP, and will be subject to an annual QAP inspection.

BKL:kj

