



STATE OF WASHINGTON  
WASHINGTON STATE PATROL  
FORENSIC LABORATORY SERVICES BUREAU

2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2027 • (206) 262-6000 • FAX (206) 262-6018

October 21, 2004

Sergeant Rod Gullberg  
Washington State Patrol  
Breath Test Program  
811 E Roanoke  
Seattle WA 98102

Dear Sergeant Gullberg:

Thank you for your recent letter (attached) containing the identifiers of the software to be used in the DataMaster CDM. I have reviewed the performance testing data obtained using this software and am approving that software version for further evaluation.

Tests conducted with this software will be indicated by software version 76043-004 (04/28/04) printed across the top of the breath test document.

Version C (102497.BIN) of the software previously approved on November 22, 1999 is no longer in use and is not approved for use for any tests conducted after today's date.


With this approval, as of today the currently approved software versions are:

DataMaster  
Version D: 072299.BIN

DataMaster CDM  
76043-004 (4/28/04) 042804.OBJ

Thank you for your continued careful oversight of the breath test program.

Sincerely,

  
Barry K. Logan Ph.D. DABFT  
Washington State Toxicologist

BKL:kj  
Attachment





STATE OF WASHINGTON  
WASHINGTON STATE PATROL  
811 E. Roanoke • Seattle, Washington 98102 • (206) 720-3018

October 21, 2004

Dr. Barry K. Logan  
Washington State Toxicologist  
Park 90/5 Suite 360  
2203 Airport Way South  
Seattle, Washington 98134-2027

Dear Dr. Logan,

The Washington State Patrol Breath Test Section has recently evaluated the Datamaster CDM for the purpose of performing evidential breath alcohol analysis in DUI enforcement in Washington State. This instrument contains software largely identical to that currently approved for use in the Datamaster instruments (072299.BIN). The recent evaluation and successful performance of the Datamaster CDM included the evaluation of the software within the instrument.

The version of software used in the Datamaster CDM is identified as version number 76043-004 and dated 04/28/04. The version number and date will be displayed on all printout documents. The computer file name identifying the new software is 042804.OBJ.

Valid performance of this software will be ensured through instrumental compliance with the Quality Assurance Procedure and the standards associated with a field evidential test as outlined in the Washington Administrative Code (WAC 448-16).

I am requesting your formal approval of this BAC Datamaster CDM software version 76043-004 04/28/04.

Sincerely,

CHIEF LOWELL PORTER

Sgt. Rod G. Gullberg  
Breath Alcohol Test Section

RGG:rg

