



COVIDIEN

CERTIFICATE OF COMPLIANCE


This is to certify that the products listed below are in compliance with the current FDA Quality System Requirements as stipulated in 21 CFR Part 820. Representative product was inspected and tested in accordance with current Kendall specifications and quality requirements.

<u>Product Code</u>	<u>Description</u>	<u>Lot Number</u>
8881352796	TT-BCS GRA 16x100 10ML P.O.-F	327043

This tube was designed for laboratory procedures requiring plasma or whole blood and chemistry procedures where glycoltic inhibition of the specimen is required. The tube was manufactured to the following specifications:

	Range	Nominal
Potassium Oxalate, ACS:	15.0 – 26.0 mg	20 mg
Sodium Fluoride, ACS:	88.0 – 112.0 mg	99 mg
Draw Volume:	9.0 -11.0 mL	10 ml

*[Signature]* 10/23/13  
 Tom Schmidt  
 Quality Assurance Manager  
 Covidien, Norfolk NE



*[Signature]*  
 10.23.13

RECEIVED

OCT 24 2013

WSP TOX LAB