

**Kendall**

[www.tycohealthcare.com](http://www.tycohealthcare.com)

To: Ms. Jennifer Rice  
Washington State Toxicology Lab  
2203 Airport Way S. Ste 360  
Seattle, WA 98134

Fax: 206-262-6145

RECEIVED  
FEB 26 2009  
WSP TOX LAB

**CERTIFICATE OF COMPLIANCE**

This is to certify that the products listed below are in compliance with the current FDA Quality System Requirements (QSR) as stipulated in 21 CFR Part 820. Representative product was inspected and tested in accordance with current Kendall specifications and quality requirements.

<u>Product #</u>	<u>Description</u>	<u>Lot #</u>
8881352788	BCS GRA 16X100 10ML P.O+F	832118

This tube was designed for laboratory procedures requiring plasma or whole blood and chemistry procedures where glycolytic inhibition of the specimen is required. The tube was manufactured to the following specifications:

	Range	Nominal
Potassium Oxalate	15.0-26.0mg	20mg
Sodium Fluoride	20.0-32.0mg	25mg
Draw volume	9.0-11.0ml	10ml

David A. Olson

David A. Olson  
Vice President Regulatory Affairs

On this 19 day of February, 2009, before me the undersigned notary public, personally appeared DAVID OLSON, proved to me through satisfactory evidence of identification, which were personal knowledge to be the person whose name is signed on the proceeding or attached document in my presence.

Dawn M Pawlitschek

Notary Public

My commission expires:

~~Dawn M. Pawlitschek~~  
NOTARY PUBLIC  
My commission expires Apr. 17, 2009

Date: February 9, 2009