

Kendall

To: Ileah Kang
Washington State Toxicology Lab
2203 Airport Way S. Ste 360
Seattle, WA 98134

Fax: 206-262-6145

RECEIVED
FEB 17 2010
WSP TOX LAB

CERTIFICATE OF COMPLIANCE

This is to certify that the products listed below are in compliance with the current FDA Quality System Requirements (QSR) as stipulated in 21 CFR Part 820. Representative product was inspected and tested in accordance with current Kendall specifications and quality requirements.

<u>Product #</u>	<u>Description</u>	<u>Lot #</u>
8881352788	BCS GRA 16X100 10ML P.O+F	922607

This tube was designed for laboratory procedures requiring plasma or whole blood and chemistry procedures where glycoltic inhibition of the specimen is required. The tube was manufactured to the following specifications:

	Range	Nominal
Potassium Oxalate	15.0-26.0mg	20mg
Sodium Fluoride	20.0-32.0mg	25mg
Draw volume	9.0-11.0ml	10ml

David A. Olson
David A. Olson
Vice President Regulatory Affairs

On this 11th day of February, 2010, before me the undersigned notary public, personally appeared David A. Olson, proved to me through satisfactory evidence of identification, which were personally known to me to be the person whose name is signed on the proceeding or attached document in my presence.

Nancy E. Paglin
Notary Public
My commission expires: November 28, 2014

Date: January 25, 2010