



STATE OF WASHINGTON
DEPARTMENT OF AGRICULTURE
Weights and Measures Program Metrology Laboratory
2747 29th Avenue Southwest • Tumwater, Washington 98512-6104
Ph (360) 753-5043 • Fax (360) 586-4728 • e-mail dwright@agr.wa.gov

REPORT OF CALIBRATION

Issued To:

Washington State Patrol
8623 Armstrong Road SW
Olympia, WA 98504

Point of Contact:

Dave Cromer
Ph. 360-596-6000

Purchase Order Number:


N/A

Report Number:


200446-0-L2902-1

Calibration Date: November 17, 2009

This is to certify that the information contained in this report is true and correct as of the date of calibration.



Dan Wright, State Metrologist



Date of Issue



Accredited by the National Laboratory Accreditation Program for the specific scope of accreditation under lab code 200446-0. This report may not be used to claim product endorsement by NVLAP or any other government agency, and may not be reproduced, except in full, without written approval from the laboratory.

WSDA Weights and Measures Metrology Laboratory

Report Number: 200446-0-L2902-1

Calibration Date: November 17, 2009

Artifact(s) Description

Test Item: Weight Cart
Serial Number: 105-6-93
Manufacture: Weight Carts, Inc.
Condition: Good
Date Received: November 16, 2009
Model Number: WC-20K
Tolerance Specification: NIST HB 105-8
Seal Number: 0262448

Calibration Information

Job Order #: L2902
Metrologist: Dan Wright
Procedure: NIST IR 6969, SOP 4
Temperature: 18.8 °C
Pressure: 752.3 mm Hg
Humidity: 41.9 % RH

Laboratory Reference Standards Used

Description	Serial Number	Cert. Number	Cal Date	Cal Due
1000 lb - 10 lb	SET WC	L2781-1	11/18/2008	11/18/2010

Traceability Statement

The artifact(s) described in this report have been compared to the Standards of the State of Washington. The Standards of the State of Washington are traceable to the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The report number for this report is the only unique report number to be used in referencing measurement traceability for the artifact(s) described in this report.

Uncertainty Statement

The combined standard uncertainty includes uncertainties reported for the standard, uncertainties associated with the measurement process, uncertainties for any observed deviations from reference values which are less than surveillance limits, and other uncertainties associated with the particular artifact (i.e., material density, air buoyancy corrections, etc.). The combined standard uncertainty is multiplied by k , a coverage factor of 2, to give the expanded uncertainty (which defines an interval with an approximate 95 percent level of confidence). The expanded uncertainty presented in this report is consistent with NIST Technical Note 1297. Stated uncertainties are less than 1/3 of the applicable tolerances. Magnetic testing has not been performed and there are no components for the effects of magnetism in the uncertainty budget.

Certification Statement

Accredited by the National Institute of Standards and Technology (NIST) through the National Voluntary Laboratory Accreditation Program (NVLAP) for the specified scope of accreditation under lab code 200446-0. This laboratory meets the requirements of ISO/IEC 17025 and ANSI/NCCL Z540-1.

WSDA Weights and Measures Metrology Laboratory

Report Number: 200446-0-L2902-1

Calibration Date: November 17, 2009

Pertinent Information

- In-accordance-with ISO/IEC FDIS 17025, General Requirements for the Competence of Testing and Calibration Laboratories, paragraph 5.10.4.4 'A calibration certificate (or calibration label) shall not contain any recommendation on the calibration interval except where this has been agreed with the client. This requirement may be superseded by legal regulations.'
- In-accordance-with Washington Administrative Code (WAC) Chapter 16-663, Service Agents -- Reporting, Test Procedures, Standards And Calibration Of Weighing And Measuring Devices, Section 16-663-130, Adequacy of standards and submission of standards for certification, paragraph 2, '... All standards used for servicing, repairing and/or calibrating commercial weighing and measuring devices must be submitted at least every two years for examination and certification...'
- Liquid levels, as stated on the attached Weight Cart Inspection Checklist, must be maintained as close to reference levels as possible during use.
- The attached Weight Cart Inspection Checklist is an integral component of this Report of Calibration and a copy must be maintained with the cart and reviewed prior to use.
- Any maintenance, repairs, replacement of parts, or damage to the weight cart or its components will likely result in an out-of-tolerance condition. Maintenance or replacement of components such as batteries, tires, filters, or other items listed on the attached Weight Cart Inspection Checklist will require calibration of the weight cart prior to subsequent use.
- The artifact(s) listed above have been found and/or left within the tolerances for the specification stated above, except as noted. An artifact is considered in-tolerance when the correction plus the measurement uncertainty is equal to or less than the specified tolerance. ***Bold Italic*** print indicates an out-of-tolerance reading.
- All corrections stated in this report correlate to a "Conventional Mass" (CM), also known as 'apparent mass', scale versus 8.0 g/cm^3 reference mass density and an air density of 1.2 mg/cm^3 at $20 \text{ }^\circ\text{C}$.
- The results listed in this report relate only to the artifacts described and extent of calibrations performed.

WSDA Weights and Measures Metrology Laboratory

Report Number: 200446-0-L2902-1

Calibration Date: November 17, 2009

As Found Calibration Results

Nominal Mass (lb)	ID	True Mass (lb)	True Mass Correction (lb)	Conventional Mass (lb)	Conventional Mass Correction (lb)	Uncertainty k=2 ± (lb)	NIST HB 105-8 Tolerance ± (lb)	Assumed Density (g/cm ³)	Air Density (g/cm ³)
4000	105-6-3	---	--	3,998.11	-1.89	0.13	1.25	7.84	0.001176505

As Left Calibration Results

Nominal Mass (lb)	ID	True Mass (lb)	True Mass Correction (lb)	Conventional Mass (lb)	Conventional Mass Correction (lb)	Uncertainty k=2 (lb)	NIST HB 105-8 Tolerance ± (lb)	Assumed Density (g/cm ³)	Air Density (g/cm ³)
4000	105-6-93	---	---	3,999.74	-0.26	0.13	1.25	7.84	0.001176505



Metrology Laboratory Form
WAMRF-011 4/9/03

Weight Cart Inspection Checklist

<u>COMPANY</u> Washington State Patrol	<u>INSPECTION DATE</u> November 16, 2009	<u>STATE TEST NO.</u> 200446-0-L2902-1
<u>ADDRESS</u> 8623 Armstrong Road SW Olympia, WA 98504	<u>NOMINAL VALUE</u> 3975 lb	<u>MODEL NUMBER</u> WC-20k
	<u>MANUFACTURER</u> Weight Carts, Inc.	<u>SERIAL NUMBER</u> 105-6-93
<u>POINT OF CONTACT</u> Greg Parriott	<u>PHONE NUMBER</u> 360-596-6000	

Power: Electric Battery ❖ Electric Generator ❖ Gasoline ❖ Diesel

Fluids:	<input type="checkbox"/> Engine Oil		Reference Level:
	<input type="checkbox"/> Hydraulic Oil	Sealed? Yes <input type="checkbox"/> No <input type="checkbox"/>	Reference Level:
	<input checked="" type="checkbox"/> Battery	Sealed? Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>	Reference Level: N/A
	<input type="checkbox"/> Fuel	Sight Gauge? Yes <input type="checkbox"/> No <input type="checkbox"/>	Reference Level:

Number of axles: 2

Number / Size of Tires: 2 @ 18 X 7 X 12 1/2 and 2 @ 21 X 7 X 15

Nominal mass of weight cart is suitably marked. Yes No

Do fluid drain tubes extend beyond the body of the cart? Yes No N/A

Sealed wheel bearings. Yes No

Drain holes present in locations where water may accumulate. Yes No

Weight restraint railing permanently fixed and solid. Yes No

Adjusting cavity is accessible. Yes No Approximate capacity: 20 lb

Adjusting cavity sealed. Yes No

Service brakes are functioning properly. Yes No

Parking brakes are functioning properly. Yes No

Remote control is functioning properly. Yes No N/A

General condition at time of calibration (i.e., any accumulated dirt/debris, damage, loose parts, or evidence of tampering or unauthorized entry of seals): **Good**

List and report any repair and/or maintenance performed (i.e., leaks repaired, parts replaced, wheels changed, welding performed, etc.) since the last calibration: **Painted**

Authorized Signature: *Dan Wright*