



PYROMETRIC COMPANY

1312 South 96TH Street
Seattle, Washington 98108
(206) 762-8307 www.pyrometric.com Fax (206) 763-9459


CALIBRATION REPORT

Customer: WASHINGTON STATE PATROL P.O.# X 00124653 Test No. 25590
Location: PROPERTY MANAGEMENT CONTROL Attn: GREG PARRIOTT Test Date: 11/13/06
Instrument: TOLEDO LOAD CELL C2P1-50K/RICE LAKE INDICATOR IQ PLUS 310 Recall Date: 11/13/08
Serial No. 94298/91441410 Range: 25,000 LBF. X 10 LBF. Req'd Accy: ± 1% Indicated

As Found Condition: In Tol. X Out of Tol. _____ Repair _____

STANDARD LBF.	INSTRUMENT 0 LBF.	CORRECTION #VALUE! LBF.
2499.7	2500	-0.3
4999.9	5000	-0.1
7499.2	7500	-0.8
9998.4	10000	-1.6
12498.1	12500	-1.9
14997.1	15000	-2.9
17496.4	17500	-3.6
19996.0	20000	-4.0
22496.0	22500	-4.0
24995.8	25000	-4.2

NOTE: READINGS OBTAINED FROM AVERAGE OF 2 RUNS



Rita E. Lingle 12-27-06
Mike Johnson 12-27-06

Condition at time of testing: Temperature: 71°F Relative Humidity: 46 %

Remarks: The above equipment meets or exceeds the requirements of the following specifications:

PER. CUSTOMER _____, _____, _____, _____, _____

Test procedure used: 4.3.2 . Our calibration system complies with ANSI/NCSL Z540-1.

All test equipment used is traceable to standards of the National Institute of Standards and Technology.

Test Equipment	Serial No.	Test Date	Recall Date	Test No.
REVERE LOAD CELL	992622	04/04/06	04/04/08	SJT.01/108944
INTERFACE LOAD CELL	117123A	04/05/06	04/05/08	SJT.01/108944

Service Technician: Mike Johnson