



WASHINGTON STATE PATROL

SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: Feb 27 2012 INITIALS: Dec OWNER: WASHINGTON STATE PATROL
 DISTRICT: 8 ADDRESS: SR101 MP77
 LOCATION Artic 77 COUNTY: Grays Harbor
 MAKE: Rice Lake TYPE Trk Scale ACCURACY CLASS: IIII
 MODEL: IQ-310 ID. NO: 98438492 CAPACITY: 100000 lbs. SECTIONS: 2
 MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM 8x10


LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
SEC. 1	20000	20000	0				
CENTER	20000	20000	0				
SEC. 2	20000	20000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L3184-2 and 200446-0-L3184-4

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: Mar 21, 2012

Print name: David Cromer



State of Washington
 County of Thurston
 Signed and sworn to (or affirmed before me on Mar. 21, 2012)

Patricia A. Geppert
 (Signature of notary public)
PATRICIA A. GEPPERT
 Printed Name of Notary
 My appointment expires Sept. 4, 2015