



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: Aug 25, 2010 INITIALS: Dec OWNER: SNOHOMISH COUNTY SHERIFF'S OFFICE
 DISTRICT: ADDRESS:
 LOCATION: COUNTY:
 MAKE: GEC TYPE PORT. WHEEL SCALE ACCURACY CLASS: IIII
 MODEL: MD-400 ID. NO: 19205 CAPACITY: 20000 lbs. SECTIONS:
 MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: DIAL PLATFORM 11"X12"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
CENTER	5000	5000	0				
CENTER	10000	10000	0				
CENTER	15000	15000	0				
CENTER	20000	20100	+100				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report and Certificate of Calibration", Certification number(s): 12345-W-02

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: Sept 2, 2010

Print name: David Cromer

State of Washington
 County of Shenandoah
 Signed and sworn to (or affirmed before me on Sept 2, 2010 by

Brenda A. Biscay
 (Signature of notary public)
Brenda A. Biscay
 Printed Name of Notary
 My appointment expires 3.12.12

