



**WASHINGTON STATE PATROL**  
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: Oct 16, 2011 INITIALS: JEC OWNER: SNOHOMISH COUNTY  
 DISTRICT: ADDRESS: COUNTY:  
 LOCATION: COUNTY:  
 MAKE: GEC TYPE PORT. WHEEL SCALE ACCURACY CLASS: IIII  
 MODEL: MD-400 ID. NO: 19205 CAPACITY: 20000 lbs. SECTIONS:  
 MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: DIAL PLATFORM 10"X11"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
CENTER	5000	5000	0				
CENTER	10000	10000	0				
CENTER	15000	14960	-40				
CENTER	20000	20000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report and Certificate of Calibration", Certification number(s): 36293-W-01

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy

Signed by: David Cromer Date: Nov 30, 2011

Print name: DAVID CROMER

State of Washington  
 County of Thurston  
 Signed and sworn to (or affirmed before me on 11-30-11 by

Patricia A. Geppert  
 (Signature of notary public)  
PATRICIA A. Geppert  
 Printed Name of Notary  
 My appointment expires Sept. 4, 2015

