



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: Aug 30, 2011 INITIALS: Dec OWNER: WASHINGTON STATE PATROL

DISTRICT: 4 ADDRESS: I-90E MP300

LOCATION Spokane 64 -N COUNTY: Spokane

MAKE: GSE TYPE Trk Scale ACCURACY CLASS: III L

MODEL: 662 ID. NO: 200345 CAPACITY: 100000 lbs. SECTIONS: 2

MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM 12X22

North

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
Platform 1				Plat 2			
Sec 1	20000	20000	0	Sec 1	20000	20000	0
Center	20000	20000	0	Center	20000	20000	0
Sec 2	20000	20000	0	Sec 2	20000	20000	0

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-1 and 200446-0-L2902-3

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: Oct 6, 2011

Print name: David Cromer

State of Washington
 County of Thurston
 Signed and sworn to (or affirmed before
 me on October 6, 2011 by

Charice L Pidcock

(Signature of notary public)

Charice L Pidcock

Printed Name of Notary

My appointment expires 12/4/14

