



WASHINGTON STATE PATROL

SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: Aug 9, 2011 INITIALS: Jec OWNER: WASHINGTON STATE PATROL
 DISTRICT: 1 ADDRESS: I-5 MP117
 LOCATION FtLewis 7 W COUNTY: Pierce
 MAKE: GSE TYPE Trk Scale ACCURACY CLASS: IIIL
 MODEL: 651 ID. NO: 200732 CAPACITY: 100000 lbs. SECTIONS: 6
 MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM 12'x32'

				TRIPLE WEST			
LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
PLAT.1 SEC.1	20000	20020	+20	PLAT.3 SEC.1	20000	20020	+20
PLAT.1 SEC.2	20000	20000	0	PLAT.3 SEC.2	20000	20000	0
PLAT.2 SEC.1	20000	20000	0				
PLAT.2 SEC.2	20000	20000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-2 and 200446-0-L2902-8

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: Oct 6, 2011

Print name: David Cromer

State of Washington
 County of Thurston
 Signed and sworn to (or affirmed before
 me on October 6, 2011 by

Charice L Pidcock
 (Signature of notary public)

Printed Name of Notary Charice L Pidcock
 My appointment expires 12/4/14

