WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: July 20, 2010 INITIALS: DEC OWNER: WASHINGTON STATE PATROL

DISTRICT: 7 ADDRESS: SR537 MP57

LOCATION: Anacortes 35 COUNTY: Skagit

MAKE: GSE TYPE Trk Scale ACCURACY CLASS: IIIL

MODEL: 651 ID. NO: 201471 CAPACITY: 100000 lbs. SECTIONS: 2

MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM 12x20 Dual

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
PLATFORM 1				PLATFORM 2			
SEC 1	20000	20000	0	SEC 1	20000	20000	0
CENTER	20000	20000	0	CENTER	20000	20000	0
SEC 2	20000	20000	0	SEC 2	20000	20000	0

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-2 and 200446-0-L2902-8

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: Aug 4, 2010

Print name: David Cromer

State of Washington
County of Thurston
Signed and sworn to (or affirmed before me on Aug 4, 2010 by
Brenda A Biscay
(Signature of notary public)
Brenda A Biscay
Printed Name of Notary
My appointment expires 3.12.12

BRENDA A. BISCAY
COMMISSION EXPIRES
NOTARY
PUBLIC
MARCH 12, 2012
STATE OF WASHINGTON