



# WASHINGTON STATE PATROL

## SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: June 22, 2011 INITIALS: Dec OWNER: WASHINGTON STATE PATROL  
 DISTRICT: 6 ADDRESS: I-90W Bullfrog  
 LOCATION CleElum 53 WB S COUNTY: Kittitas  
 MAKE: GSE TYPE Trk Scale ACCURACY CLASS: IIII  
 MODEL: 651 ID. NO: 202173 CAPACITY: 200000 lbs. SECTIONS: 2  
 MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM 12x20  
Dual South

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
PLATFORM 1				PLATFORM 2			
SEC 1	20000	19980	-20	SEC 1	20000	20000	0
CENTER	20000	20000	0	CENTER	20000	20000	0
SEC 2	20000	20000	0	SEC 2	20000	20000	0

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-1 and 200446-0-L2902-3

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: July 19, 2011  
 Print name: David Cromer

State of Washington  
 County of Thurston  
 Signed and sworn to (or affirmed before me on July 19, 2011 by Charice L. Pidcock  
 (Signature of notary public)  
 Printed Name of Notary Charice L. Pidcock  
 My appointment expires 12/4/14

