



**WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION**

DATE: June 1, 2010 INITIALS: DCC OWNER: WASHINGTON STATE PATROL
 DISTRICT: 6 ADDRESS: I-90W Bullfrog
 LOCATION CleElum 53 WB N COUNTY: Kittitas
 MAKE: GSE TYPE Trk Scale ACCURACY CLASS: III L
 MODEL: 651 ID. NO: 202174 CAPACITY: 200000 lbs. SECTIONS: 2
 MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM 12x20

				Dual		North	
LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
PLATFORM 1				PLATFORM 2			
SEC 1	20000	20000	0	SEC 1	20000	20000	0
CENTER	20000	20000	0	CENTER	20000	20000	0
SEC 2	20000	20000	0	SEC 2	20000	20000	0

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-1 and 200446-0-L2902-3

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: July 1, 2010
 Print name: David Cromer

State of Washington
 County of Shelton
 Signed and sworn to (or affirmed before
 me on July 1, 2010 by
Brenda A. Biscay
 (Signature of notary public)
Brenda A. Biscay
 Printed Name of Notary
 My appointment expires 3.12.12

