



WASHINGTON STATE PATROL

SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: July 19, 2010 INITIALS: DEC OWNER: WASHINGTON STATE PATROL
 DISTRICT: 7 ADDRESS: I-5 MP215
 LOCATION: Stanwood 3 E COUNTY: Snohomish
 MAKE: GSE TYPE Trk Scale ACCURACY CLASS: III L
 MODEL: 651 ID. NO: 202184 CAPACITY: 100000 lbs. SECTIONS: 2
 MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM 12x20

				Dual East			
LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
PLATFORM 1				PLATFORM 2			
SEC 1	20000	20000	0	SEC 1	20000	20000	0
CENTER	20000	20000	0	CENTER	20000	20000	0
SEC 2	20000	20020	+20	SEC 2	20000	19980	-20

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-2 and 200446-0-L2902-8

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: Aug 4, 2010
 Print name: David Cromer

State of Washington
 County of Shurston
 Signed and sworn to (or affirmed before me on Aug 4, 2010 by

Brenda A Biscay
 (Signature of notary public)
Brenda A. Biscay
 Printed Name of Notary
 My appointment expires 3.12.12

