



**WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION**


DATE: July 23 2012 INITIALS: Dec OWNER: WASHINGTON STATE PATROL
 DISTRICT: 1 ADDRESS: I-5 MP117
 LOCATION: FtLewis 7 E COUNTY: Pierce
 MAKE: GSE TYPE: Trk Scale ACCURACY CLASS: IIIL
 MODEL: 662 ID. NO: 206305 CAPACITY: 100000 lbs. SECTIONS: 4
 MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM 12'X22'

				DUAL		EAST	
LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
PLAT. 1 SEC.1	20000	20000	0	PLAT.2 CENT.	20000	20000	0
PLAT. 1 CENT.	20000	20000	0	PLAT.2 SEC.2	20000	20000	0
PLAT.1 SEC.2	20000	20000	0				
PLAT.2 SEC.1	20000	20000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L3184-2 and 200446-0-L3184-4

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: Aug 20, 2012
 Print name: David Cromer



State of Washington
 County of Thurston
 Signed and sworn to (or affirmed before me on August 20, 2012

Patricia A. Geppert
 (Signature of notary public)
PATRICIA A. Geppert

Printed Name of Notary
 My appointment expires Sept. 4, 2015