



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: MARCH 14, 2011 INITIALS: DJL OWNER: WASHINGTON STATE PATROL

DISTRICT: 7 ADDRESS:

LOCATION: FM SHOP

COUNTY:

MAKE: HAENNI

TYPE PORT. WHEEL SCALE

ACCURACY CLASS: IIII

MODEL: 101

ID. NO: 21809

CAPACITY: 20000 lbs. SECTIONS:

MINIMUM GRADUATIONS: 50 lbs.

REG. ELEMENT: DIAL

PLATFORM 15"X26"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
CENTER	5000	5000	0				
CENTER	10000	10000	0				
CENTER	15000	15000	0				
CENTER	20000	19950	-50				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report and Certificate of Calibration", Certification number(s): 36293-W-02

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: Dan Lafreniere

Date: March 28, 2011

Print name: DAN LAFRENIERE

State of Washington

County of Shelton

Signed and sworn to (or affirmed before me on Mar. 28, 2011 by

Brenda A. Biscay

(Signature of notary public)

Brenda A. Biscay

Printed Name of Notary

My appointment expires 3.12.12

