WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: Dec. 21, 2011 INITIALS: DJL OWNER: CITY OF TACOMA

DISTRICT: ADDRESS:

LOCATION: COUNTY:

MAKE: HAENNI TYPE PORT. WHEEL SCALE ACCURACY CLASS: IIII

MODEL: 101 ID. NO: 21955 CAPACITY: 20000 lbs. SECTIONS:

MINIMUM GRADUATIONS: 50 lbs. REG. ELEMENT: DIAL PLATFORM 15"X26"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
CENTER	5000	5000	0				
CENTER	10000	10000	0				
CENTER	15000	15000	0				
CENTER	20000	19950	-50				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report and Certificate of Calibration", Certification number(s): 36293-W-01

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: Dan LaFreniere Date: 12/21/2011

Print name: DAN LAFRENIERE

State of Washington
County of Thurston
Signed and sworn to (or affirmed before me on Dec. 21, 2011 by
Patricia A. Geppert
(Signature of notary public)
PATRICIA A. GEPPERT
Printed Name of Notary
My appointment expires Sept. 4, 2015

PATRICIA A. GEPPERT
COMMISSION EXPIRES
NOTARY
PUBLIC
SEPT. 4, 2015
STATE OF WASHINGTON