



WASHINGTON STATE PATROL  
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: DEC. 6, 2010 INITIALS: DJL OWNER: CITY OF TACOMA  
DISTRICT: ADDRESS: COUNTY:  
LOCATION: TYPE PORT. WHEEL SCALE ACCURACY CLASS: IIII  
MAKE: HAENNI ID. NO: 21966 CAPACITY: 20000 lbs. SECTIONS:  
MODEL: 101 REG. ELEMENT: DIAL PLATFORM 15"X26"  
MINIMUM GRADUATIONS: 50 lbs.

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
CENTER	5000	5000	0				
CENTER	10000	10000	0				
CENTER	15000	15000	0				
CENTER	20000	20050	+50				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report and Certificate of Calibration", Certification number(s): 36293-W-02

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy

Signed by: Dan LaFreniere Date: 1/20/2011  
Print name: Dan LaFreniere

State of Washington  
County of Shurston  
Signed and sworn to (or affirmed before me on Jan 20, 2011 by

Brenda A Biscay  
(Signature of notary public)  
Brenda A. Biscay  
Printed Name of Notary  
My appointment expires 3.12.12

