



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: Jan 26, 2012 INITIALS: DEC OWNER: WASHINGTON STATE PATROL
DISTRICT: 3 ADDRESS:
LOCATION: FM SHOP COUNTY:
MAKE: HAENNI TYPE PORT. WHEEL SCALE ACCURACY CLASS: IIII
MODEL: 101 ID. NO: 23081 CAPACITY: 20000 lbs. SECTIONS:
MINIMUM GRADUATIONS: 50 lbs. REG. ELEMENT: DIAL PLATFORM 15"X26"

Table with 8 columns: LOAD POSITION, TEST WEIGHT, SCALE INDICATION, ERROR, LOAD POSITION, TEST WEIGHT, SCALE INDICATION, ERROR. Rows show test results for 5000, 10000, 15000, and 20000 lbs weights, all with 0 error.

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report and Certificate of Calibration", Certification number(s): 36293-W-01

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: Feb 21, 2012
Print name: DAVID CROMER



State of Washington
County of Thurston
Signed and sworn to (or affirmed before me on Feb. 21, 2012 by Patricia A. Geppert
(Printed Name of Notary)
My appointment expires Sept. 4, 2015