



**WASHINGTON STATE PATROL**  
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: Nov 30, 2011 INITIALS: DEC OWNER: CITY OF SEATTLE  
 DISTRICT: ADDRESS:  
 LOCATION: COUNTY:  
 MAKE: INTERCOMP TYPE PORT. WHEEL SCALE ACCURACY CLASS: IIII  
 MODEL: PT-300 ID. NO: 23083510 CAPACITY: 20000 lbs. SECTIONS:  
 MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: DIGITAL PLATFORM 10"X12"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
CENTER	5000	5000	0				
CENTER	10000	9960	-40				
CENTER	15000	14960	-40				
CENTER	20000	19940	-60				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report and Certificate of Calibration", Certification number(s): 36293-W-01

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: Dec 15, 2011

Print name: DAVID CROMER

State of Washington  
 County of Thurston  
 Signed and sworn to (or affirmed before  
 me on December 15, 2011

Patricia A. Seppert  
 (Signature of notary public)

PATRICIA A. SEPPERT  
 Printed Name of Notary  
 My appointment expires Sept. 4, 2015

