



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: May 13, 2012 INITIALS: DJK OWNER: THURSTON COUNTY SHERIFF
DISTRICT: ADDRESS:
LOCATION: FM SHOP COUNTY:
MAKE: HAENNI TYPE: PORT WHEEL SCALE ACCURACY CLASS: IIII
MODEL: 101 ID. NO: 23170 CAPACITY: 20000 lbs. SECTIONS:
MINIMUM GRADUATIONS: 50 lbs. REG. ELEMENT: DIAL PLATFORM 15"X26"


LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
CENTER	5000	5000	0				
CENTER	10000	10000	0				
CENTER	15000	15000	0				
CENTER	20000	19950	-50				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report and Certificate of Calibration", Certification number(s): 36293-W-02

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: Dan Lafreniere Date: May 14, 2012

Print name: DAN LAFRENIERE



State of Washington
County of Thurston
Signed and sworn to (or affirmed before me on May 14, 2012
Patricia A. Geppert
(Signature of notary public)
PATRICIA A. Geppert
Printed Name of Notary
My appointment expires Sept. 4, 2015