



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: May 13, 2012 INITIALS: DJL OWNER: THURSTON COUNTY SHERIFF
 DISTRICT: _____ ADDRESS: _____
 LOCATION: FM SHOP COUNTY: _____
 MAKE: HAENNI TYPE: PORT. WHEEL SCALE ACCURACY CLASS: IIII
 MODEL: 101 ID. NO: 23172 CAPACITY: 20000 lbs. SECTIONS:
 MINIMUM GRADUATIONS: 50 lbs. REG. ELEMENT: DIAL PLATFORM 15"X26"


LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
CENTER	5000	5000	0				
CENTER	10000	10000	0				
CENTER	15000	15000	0				
CENTER	20000	19950	-50				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report and Certificate of Calibration", Certification number(s): 36293-W-02

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: Dan Lafreniere Date: May 14, 2012

Print name: DAN LAFRENIERE



State of Washington
 County of Thurston
 Signed and sworn to (or affirmed before me on May 14, 2012)
Patricia A. Geppert
 (Signature of notary public)
PATRICIA A. Geppert
 Printed Name of Notary
 My appointment expires Sept. 4, 2015