



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: April 4, 2011 INITIALS: DC OWNER: WASHINGTON STATE PATROL
 DISTRICT: 1 ADDRESS:
 LOCATION: FM SHOP COUNTY:
 MAKE: HAENNI TYPE PORT. WHEEL SCALE ACCURACY CLASS: IIII
 MODEL: 101 ID. NO: 24085 CAPACITY: 20000 lbs. SECTIONS:
 MINIMUM GRADUATIONS: 50 lbs. REG. ELEMENT: DIAL PLATFORM 15"X26"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
CENTER	5000	5000	0				
CENTER	10000	10000	0				
CENTER	15000	15000	0				
CENTER	20000	19950	-50				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report and Certificate of Calibration", Certification number(s): 36293-W-01

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: April 13, 2011
 Print name: DAVID CROMER

State of Washington
 County of Thurston
 Signed and sworn to (or affirmed before me on April 13, 2011 by
Charice L. Pidcock
 (Signature of notary public)
Charice L. Pidcock
 Printed Name of Notary
 My appointment expires 12/4/14

