



**WASHINGTON STATE PATROL**  
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: Aug 9, 2010 INITIALS: Jec OWNER: WASHINGTON STATE PATROL  
 DISTRICT: 8 ADDRESS: SR101W 254  
 LOCATION Pt Angeles 15 COUNTY: Clallum  
 MAKE: Weightronix TYPE Trk Scale ACCURACY CLASS: III  
 MODEL: W1110 ID. NO: 26481 CAPACITY: 100000 lbs. SECTIONS: 2  
 MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Electronic PLATFORM 10x12

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
SEC. 1	20000	20000	0				
CENTER	20000	20000	0				
SEC. 2	20000	20000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-2 and 200446-0-L2902-8

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: Aug 24, 2010  
 Print name: David Cromer

State of Washington  
 County of Shenston  
 Signed and sworn to (or affirmed before  
 me on Aug 24, 2010 by  
Brenda A Biscay  
 (Signature of notary public)  
Brenda A. Biscay  
 Printed Name of Notary  
 My appointment expires 3.12.12

