# WASHINGTON STATE PATROL
## SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: Mar 16, 2010 INITIALS: DEC OWNER: LEWIS COUNTY SHERIFF'S DEPT.

DISTRICT: ADDRESS:

LOCATION: COUNTY:

MAKE: HAENNI TYPE PORT. WHEEL SCALE ACCURACY CLASS: III

MODEL: 101 ID. NO: 26702 CAPACITY: 20000 lbs. SECTIONS:

MINIMUM GRADUATIONS: 50 lbs. REG. ELEMENT: DIAL PLATFORM 15"X26"

| LOAD POSITION | TEST WEIGHT | SCALE INDICATION | ERROR | LOAD POSITION | TEST WEIGHT | SCALE INDICATION | ERROR |
|---|---|---|---|---|---|---|---|
| CENTER | 5000 | 5000 | 0 | | | | |
| | | | | | | | |
| CENTER | 10000 | 10000 | 0 | | | | |
| | | | | | | | |
| CENTER | 15000 | 15000 | 0 | | | | |
| | | | | | | | |
| CENTER | 20000 | 20000 | 0 | | | | |
| | | | | | | | |

**Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Certificate of Accuracy", No. 12345-W-02.**

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cramer Date: Mar 23, 2010

State of Washington
County of Thurston
Signed and sworn to (or affirmed before me on Mar 23, 2010 by
Brenda A Biscay
(Signature of notary public)
Brenda A. Biscay
Printed Name of Notary
My appointment expires 3-12-12

BRENDA A. BISCAY
COMMISSION EXPIRES
NOTARY
PUBLIC
3-12-12
STATE OF WASHINGTON

3000-362-021(R)(10/98)