



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: March 30, 2011 INITIALS: W/L OWNER: LEWIS COUNTY SHERIFF'S DEPT.

DISTRICT: ADDRESS:

LOCATION: COUNTY:

MAKE: HAENNI TYPE PORT. WHEEL SCALE ACCURACY CLASS: IIII

MODEL: 101 ID. NO: 26703 CAPACITY: 20000 lbs. SECTIONS:

MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: DIAL PLATFORM 15"X26"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
CENTER	5000	5000	0				
CENTER	10000	10000	0				
CENTER	15000	15000	0				
CENTER	20000	19950	-50				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report and Certificate of Calibration", Certification number(s): 36293-W-02

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: Dan Lafreniere Date: April 5, 2011

Print name: DAN LAFRENIERE

State of Washington
County of In Thurston
Signed and sworn to (or affirmed before
me on April 5, 2011 by

Charice L. Pidcock
(Signature of notary public)

Charice L. Pidcock
Printed Name of Notary

My appointment expires 12/4/14

