



**WASHINGTON STATE PATROL**  
**SCALE TEST REPORT AND CERTIFICATION OF INSPECTION**

DATE: April 19, 2012 INITIALS: DJL OWNER: LEWIS COUNTY SHERIFF'S OFFICE  
 DISTRICT: ADDRESS: COUNTY:  
 LOCATION: FM SHOP TYPE: PORT. WHEEL SCALE ACCURACY CLASS: IIII  
 MAKE: HAENNI ID. NO: 26705 CAPACITY: 20000 lbs. SECTIONS:  
 MODEL: 101 REG. ELEMENT: DIAL PLATFORM 15"X26"  
 MINIMUM GRADUATIONS: 50 lbs.


LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
CENTER	5000	5000	0				
CENTER	10000	10000	0				
CENTER	15000	15000	0				
CENTER	20000	20000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report and Certificate of Calibration", Certification number(s): 36293-W-02

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: Dan Lafreniere Date: April 19, 2012

Print name: DAN LAFRENIERE



State of Washington  
 County of Thurston  
 Signed and sworn to (or affirmed before me on) April 19, 2012  
Patricia A. Geppert  
 (Signature of notary public)  
Patricia A. Geppert  
 Printed Name of Notary  
 My appointment expires Sept. 4, 2015