



**WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION**

DATE: Mar 16, 2010 INITIALS: DEC OWNER: LEWIS COUNTY SHERIFF'S DEPT.
 DISTRICT: ADDRESS:
 LOCATION: COUNTY:
 MAKE: HAENNI TYPE PORT. WHEEL SCALE ACCURACY CLASS: IIII
 MODEL: 101 ID. NO: 26707 CAPACITY: 20000 lbs. SECTIONS:
 MINIMUM GRADUATIONS: 50 lbs. REG. ELEMENT: DIAL PLATFORM 15"X26"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
CENTER	5000	5000	0				
CENTER	10000	10000	0				
CENTER	15000	15000	0				
CENTER	20000	19950	-50				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Certificate of Accuracy", No. 12345-W-02.

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cronin Date: Mar 23, 2010

State of Washington
 County of Shurston
 Signed and sworn to (or affirmed before
 me on Mar 23, 2010 by
Brenda A. Biscay
 (Signature of notary public)
Brenda A. Biscay
 Printed Name of Notary
 My appointment expires 3.12.12

