



**WASHINGTON STATE PATROL  
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION**

DATE: Mar 14, 2012 INITIALS: DEC OWNER: GRAYS HARBOR COUNTY  
 DISTRICT: ADDRESS:  
 LOCATION: FM SHOP COUNTY:  
 MAKE: HAENNI TYPE PORT. WHEEL SCALE ACCURACY CLASS: IIII  
 MODEL: 101 ID. NO: 26935 CAPACITY: 20000 lbs. SECTIONS:  
 MINIMUM GRADUATIONS: 50 lbs. REG. ELEMENT: DIAL PLATFORM 15"X26"


| LOAD POSITION | TEST WEIGHT | SCALE INDICATION | ERROR | LOAD POSITION | TEST WEIGHT | SCALE INDICATION | ERROR |
|---------------|-------------|------------------|-------|---------------|-------------|------------------|-------|
| CENTER        | 5000        | 5000             | 0     |               |             |                  |       |
| CENTER        | 10000       | 10000            | 0     |               |             |                  |       |
| CENTER        | 15000       | 15050            | +50   |               |             |                  |       |
| CENTER        | 20000       | 20000            | 0     |               |             |                  |       |

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report and Certificate of Calibration", Certification number(s): 36293-W-01

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: Mar 21, 2012

Print name: DAVID CROMER



State of Washington  
 County of Thorston  
 Signed and sworn to (or affirmed before me on MAR. 21, 2012)

Patricia A. Geppert  
 (Signature of notary public)  
PATRICIA A. Geppert  
 Printed Name of Notary  
 My appointment expires Sept. 4, 2015