# WASHINGTON STATE PATROL
## SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: Mar 21, 2012 INITIALS: DEC OWNER: WASHINGTON STATE PATROL

DISTRICT: 1 ADDRESS:

LOCATION: FM SHOP COUNTY:

MAKE: HAENNI TYPE PORT. WHEEL SCALE ACCURACY CLASS: IIII

MODEL: 101 ID. NO: 31972 CAPACITY: 20000 lbs. SECTIONS:

MINIMUM GRADUATIONS: 50 lbs. REG. ELEMENT: DIAL PLATFORM 15"X26"

| LOAD POSITION | TEST WEIGHT | SCALE INDICATION | ERROR | LOAD POSITION | TEST WEIGHT | SCALE INDICATION | ERROR |
|---|---|---|---|---|---|---|---|
| CENTER | 5000 | 5000 | 0 | | | | |
| | | | | | | | |
| CENTER | 10000 | 10000 | 0 | | | | |
| | | | | | | | |
| CENTER | 15000 | 15000 | 0 | | | | |
| | | | | | | | |
| CENTER | 20000 | 20000 | 0 | | | | |
| | | | | | | | |

**Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report and Certificate of Calibration", Certification number(s): 36293-W-01**

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: Mar 22, 2012

Print name: DAVE CROMER

PATRICIA A. GEPPERT
COMMISSION EXPIRES
NOTARY
PUBLIC
SEPT. 4, 2015
STATE OF WASHINGTON

State of Washington
County of Thurston
Signed and sworn to (or affirmed before me on MAR. 22, 2012 by
Patricia A. Geppert
(Signature of notary public)
PATRICIA A. GEPPERT
Printed Name of Notary
My appointment expires Sept. 4, 2015

3000-362-021(R)(10/98)