



**WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION**

DATE: July 26, 2012 INITIALS: [Signature] OWNER: WASHINGTON STATE PATROL
 DISTRICT: 2 ADDRESS: I-5S MP141
 LOCATION: SeaTac SB 27 E1 COUNTY: King
 MAKE: Toledo TYPE: Trk Scale ACCURACY CLASS: IIII
 MODEL: 8530 ID. NO: 4337885-4NV CAPACITY: 100000 lbs. SECTIONS: 2
 MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM 12x20


East1

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
SEC 1	20000	20000	0				
CENTER	20000	20000	0				
SEC 2	20000	19980	-20				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L3184-2 and 200446-0-L3184-4

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: [Signature] Date: Aug 20, 2012
 Print name: David Cromer



State of Washington
 County of Thurston
 Signed and sworn to (or affirmed before me on Aug. 20, 2012
[Signature]
 (Signature of notary public)
PATRICIA A. GEPPERT
 Printed Name of Notary
 My appointment expires Sept. 4, 2015